

PRODUCT xm18s15 — A HYDROCOLLOID TRANSFER TAPE

FEATURES

- Very good initial adhesion
- Breathable
- Gamma Sterilizable
- Good for skin attachment applications
- Gentle adhesion on skin
- Good shear adhesion property
- Good for catheter fixation

COMPOSITION

- Release liner 1: 2.0 mil clear siliconized polyethylene release liner
- Adhesive: 15 mil hydrocolloid adhesive
- Release liner 2: 3.4 mil 60# SCK paper release liner

TYPICAL PROPERTIES

| Test | Method | Typical Average Value |
|-------------------|---------------------------|-----------------------|
| Probe Tack | Modified ASTM D2979 | 450-750 grams |
| Static Shear | Modified ASTM D3654A | >2,300 minutes |
| Saline Absorption | Xennovate Internal Method | 6-20% |
| Liner Release | ASTM D 5357 | 60-120 grams |

Test method details are available upon request. Values listed here are typical average values and should not be interpreted as a specification.

ROLL DESCRIPTION

Standard rolls are 9.25" wide, 450 feet long, 6" core.

RECOMMENDED STORAGE

Store in sealed, poly bags at normal room temperature (65 to 75 ° F) and dry environment. Avoid storage in temperatures above 85°F. Keep rolls in the end-cap suspended packaging supplied by Xennovate Medical®.

SHELF LIFE

To obtain best performance, use this product within 12 months from date of manufacture.

XENNOVATE MEDICAL®

Please contact us at +1 440-974-3080 with questions or to place orders or visit us at: Xennovate.com

LIMITATION OF LIABILITY AND WARRANTY INFORMATION

Product and Safety Information: User is solely responsible for determining the suitability of Xennovate samples and products for the intended use including any necessary safety or toxicity assessment. Product has a shelf life of two years when stored at 23°C (73°F), 50% relative humidity, out of direct sunlight and in original packaging. Xennovate will provide Material Safety Data Sheets and results of any toxicity testing upon request. In every case before use or production of any product, users should conduct their own tests to determine to their own satisfaction whether the product is of acceptable quality and is suitable for their particular purposes under appropriate operating conditions.

Notice: Nothing contained herein shall be construed to imply the nonexistence of any relevant patents or to constitute a permission, inducement or recommendations to practice any invention covered by any patent, without authority from the owners of such patent.

Warranty Information: All statements, technical information and recommendations herein are based on tests Xennovate believes to be reliable, but the accuracy or completeness thereof is not guaranteed. Xennovate warrants only that products will meet Xennovate's product specifications at the time of shipment to the customer. All claims pursuant to such warranty must be brought within the earlier of (a) the time period stated on the product specification sheet or literature, or (b) one year following the date of shipment. Upon notice of any warranty claim, Xennovate's liability shall be limited, in Xennovate's sole discretion, to suitable repair or replacement, which shall not exceed the purchase price paid for the products. Xennovate does not offer any other warranty and does not warrant the performance, safety or such other characteristics of products in combination with other materials. Xennovate specifically DOES NOT warrant products for any intended or unintended uses (whether or not foreseeable); for compatibility or suitability with other components or compatibility with any methods of manufacture or conversion. The foregoing warranty is the sole and exclusive warranty given by Xennovate, and is made in lieu of and Xennovate specifically disclaims all other warranties, expressed or implied, including the implied warranties of merchantability, fitness for a particular purpose and freedom from non-infringement. No alteration to this warrant shall be valid unless given in writing and signed by an officer of Xennovate, LLC.

Limitation of Remedies: If products are proven not to meet Xennovate's specifications, the sole and exclusive remedy available and Xennovate's only obligation shall be, at Xennovate's option, to replace such quantity of Products which are proven out of specification or to refund the purchase price paid for Products.

Limitations of Liabilities: The remedies provided herein are exclusive remedies against Xennovate for any alleged or actual nonconformance to specifications or defect or other failure in products or for Xennovate's performance of its supply obligations. Under no circumstances is Xennovate liable for any indirect, incidental, special or consequential damages (including lost profits or loss of production) in any way related to the product under any theory of law including, but not limited to, negligence and strict liability. The remedies of purchaser set forth herein are exclusive, and the total liability of Xennovate in connection with the manufacture, sale, delivery, installation, shipment or use of the product, under any theory, warranty or otherwise, shall not exceed the purchase price paid for the products upon which the claim is based. This limitation of liability and warranty shall be governed by the laws of the state of Ohio, without regard to conflict of laws provisions.

XENNOVATE, LLC D/B/A XENNOVATE MEDICAL®